

Released:	P. Quinn	
Name	Date	Signature

This page was intentionally left blank

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	3 of <a href="#">51</a>

#### Change record

Issue/Rev.	Date	Section/Parag. affected	Reason/Initiation/Documents/Remarks
0.9	01/10/2016	All	Draft version

This page was intentionally left blank

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	5 of 51

## Contents

<b>1</b>	<b>Introduction</b>	<b>9</b>
1.1	Reference and applicable documents . . . . .	9
<b>2</b>	<b>GRAVITY Instrument Description</b>	<b>11</b>
<b>3</b>	<b>Quick start</b>	<b>12</b>
3.1	The concept of recipe . . . . .	12
3.2	The concept of SOF . . . . .	12
3.3	Instrument calibrations . . . . .	12
3.4	From raw data to raw visibilities . . . . .	12
3.5	From raw visibilities to calibrated visibilities . . . . .	13
<b>4</b>	<b>Instrument Data Description</b>	<b>14</b>
4.1	RAW science data . . . . .	14
4.2	RAW calibration data . . . . .	14
4.3	STATIC calibration . . . . .	15
4.4	PRODUCT calibration data . . . . .	15
4.5	PRODUCT science data . . . . .	15
<b>5</b>	<b>Data Reduction</b>	<b>17</b>
5.1	Graphical overview of the cascade . . . . .	17
5.2	Using Gasgano . . . . .	17
5.3	Using EsoRex . . . . .	18
5.4	Using run_gravi_reduce.py python script . . . . .	18
<b>6</b>	<b>Known Problems</b>	<b>20</b>
6.1	Spectral calibration . . . . .	20
6.2	Uncertainties in products . . . . .	20
<b>A</b>	<b>Pipeline Recipes Interfaces</b>	<b>21</b>
A.1	List of all recipes . . . . .	21
A.2	gravity_dark . . . . .	21

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	6 of 51

A.3	gravity_p2vm . . . . .	22
A.4	gravity_eop . . . . .	24
A.5	gravity_vis . . . . .	25
A.6	gravity_vis_from_p2vmred . . . . .	27
A.7	gravity_viscal . . . . .	29
A.8	gravity_postprocess . . . . .	29
A.9	gravity_wavelamp . . . . .	30
A.10	gravity_disp . . . . .	31
<b>B</b>	<b>Re-creating the Static Calibration</b>	<b>33</b>
B.1	Dispersion model . . . . .	33
B.2	Earth Orientation Parameters . . . . .	33
<b>C</b>	<b>Detailed description of the data content</b>	<b>34</b>
C.1	Recommended tools to browse data . . . . .	34
C.2	Table structure common to all data . . . . .	34
C.3	P2VM product . . . . .	34
C.4	*_VIS and *_TF products . . . . .	35
C.5	SPECTRUM, PREPROC products . . . . .	35
C.6	*_P2VMRED products . . . . .	35
C.7	ASTRORED product . . . . .	39
C.8	DISP_MODEL . . . . .	39
<b>D</b>	<b>Algorithms</b>	<b>40</b>
D.1	Bias clean-up . . . . .	40
D.2	Spectrum extraction . . . . .	40
D.3	Re-interpolation to a common wavelength . . . . .	41
D.4	Extraction of the coherent fluxes and telescope fluxes via P2VM . . . . .	42
D.5	Computation of SNR . . . . .	42
D.6	Computing the vFactor . . . . .	42
D.7	Frame rejection . . . . .	43
D.8	Phase referencing . . . . .	43

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	7 of 51

D.9	Geometric flux . . . . .	44
D.10	Averaged flux estimator . . . . .	44
D.11	Averaged complex visibility estimator . . . . .	44
D.12	Averaged square visibility estimator . . . . .	45
D.13	Averaged closure-phase estimator . . . . .	45
D.14	Uncertainty on averaged quantities . . . . .	45
D.15	Processing of MET and FDDL . . . . .	46
D.16	Applying dispersion correction to MET . . . . .	47
<b>E</b>	<b>Installation</b>	<b>48</b>
E.1	Supported platforms . . . . .	48
E.2	Building the GRAVITY pipeline . . . . .	48
E.2.1	Requirements . . . . .	48
E.2.2	Compiling and installing the GRAVITY pipeline . . . . .	49
<b>F</b>	<b>Abbreviations and acronyms</b>	<b>51</b>

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	8 of <a href="#">51</a>



<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	9 of 51

## 1 Introduction

This manual is a complete description of the data reduction recipes implemented for the GRAVITY pipeline, reflecting the status of the GRAVITY pipeline as of now.

The main part of the document is focused on the main feature of the pipeline useful to the science user of GRAVITY. The detailed appendixes may be more useful to ESO staff for the purpose of long term re-calibration and data quality control.

### 1.1 Reference and applicable documents

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	10 of 51

- [1] *OIFits Explorer*. <http://www.jmmc.fr/oifitsexplorer>. 34
- [2] *QFitsView*. <http://www.mpe.mpg.de/ott/QFitsView/>. 34
- [3] ESO/SDD/DFS, <http://www.eso.org/cpl/>. *CPL home page*. 48
- [4] T. A. Pauls, J. S. Young, W. D. Cotton, and J. D. Monnier. A Data Exchange Standard for Optical (Visible/IR) Interferometry. *PASP*, 117:1255–1262, November 2005. 16, 34, 35, 51

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	11 of 51

## 2 GRAVITY Instrument Description

The GRAVITY instrument has been developed under ESO contract by the GRAVITY consortium.

The main purpose of the instrument is to measure the angular separation between stars in the vicinity of the Galactic Center black hole, and even the position and motion of the infrared emission of the black hole itself during flares.

GRAVITY recombines the light beams collected by either the four 8.2m Unit Telescopes or the four VLTI Auxiliary Telescopes. The general principle of the instrument is to recombine the light coming from one (single field mode) or two (dual field mode) astronomical light sources in two different beam combiners: the Fringe Tracker (FT) and the Science Combiner (SC). In single field mode, the light from the target is split between the FT and SC channels using a beam splitter, while in dual field mode a mirror directs the light from each object to their respective beam combiners. The FT is optimized to record fringes at very high frequency (up to 1 kHz), in order to measure and compensate in real time the atmospheric piston effect using a dedicated actuator in the instrument. As the observed objects are both within the atmospheric isoplanetic patch, the correction of the atmospheric piston by the FT stabilizes the fringes of the SC channel. This gives the possibility to integrate for up to several tens of seconds, and therefore reach a high sensitivity, even at relatively high spectral dispersion, on the SC channel.

The properties of the interference fringes are measured separately in the FT and SC beam combiners. GRAVITY measures the classical interferometric observables of any source, as the previous VLTI instruments (VINCI, MIDI, AMBER and PIONIER). The FT spectral resolution is limited to 5 spectral channels over the  $K$  band. The SC has three available spectral resolutions: low (40), medium (400) and high (4000), providing approximately 10, 200 and 1800 spectral channels over the  $K$  band.

The phases of the SC beam combiner are referenced to the FT using a metrology system that encompasses the optics of the VLTI up to the secondary mirror of the telescopes. Thanks to this link between the two beam combiners, GRAVITY provides very accurate measurements of the differential position of the fringe pattern obtained, for each baseline, between a reference star (in the FT channel) and the target star (in the SC channel). Given the VLTI FOV, the angular separation between the reference object and the science target is limited to 5 arcsec with the ATs and 2 arcsec with the UTs. Within this restricted separation, the final accuracy on the relative astrometry is expected to be of a few tens of  $\mu$ arcsec.

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	12 of 51

## 3 Quick start

### 3.1 The concept of recipe

The pipeline is built over “recipes” that implements the basic steps of the reduction. The recipe input is a Set Of File (SOF), which contains the list of files to reduce, associated with the required calibrations. Each recipe has a list of possible option. Each recipe produces one or several products. The recipes can be triggered by several mechanism:

**gasgano** is the ESO user interface to classify, select files (= construct a SOF) and launch recipes.

**esorex** is the ESO command line tool to launch a recipe on any SOF manually written by the user.

**reflex** is the ESO environment allowing to reduce an entire directory in batch mode.

The GRAVITY consortium has also built a dedicated python script which brows the content of the current directory, classifies the files, writes the SOFs and automatically triggers the corresponding recipes via esorex.

### 3.2 The concept of SOF

A SOF is a consistent list of file to be sent to a recipe. Each file in the list shall be assigned a *DO Category*, which tells the recipe its purpose in the reduction. In the GRAVITY pipeline, the DO categories are trivially built from the DPR.TYPE, DPR.CATG and PRO.CATG keywords in the FITS header.

### 3.3 Instrument calibrations

In order to reduce interferometric observations, it is mandatory to calibrate the detectors and the combiners. These calibrations are obtained via dedicated observations of the internal source with all shutters closed, one shutter open at a time, two shutters open at a time, and all shutters open.

The recipes **gravity\_dark** creates the DARK calibration product, which contains the mean detector bias and the detector readout noise. It shall match the detector and the optical setup of the observation.

The recipe **gravity\_p2vm** creates the BAD (bad pixel), FLAT (internal transmission), P2VM (internal phase and contrast) and WAVE (wavelength map) calibration product. They shall match the optical setup of the observation.

These products are all needed in order to reduce the interferometric observations of science target.

### 3.4 From raw data to raw visibilities

The first step is to reduce the raw interferometric observations into uncalibrated measurement of the visibilities and closure phases. This step is generally done file-per-file, that is each OBJECT exposure in the raw directory- has a corresponding files in product directory (although it is possible to reduce several files together). Note that

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	13 of 51

this step includes the averaging over all the frames that compose a standard exposure. This step is rather long (several hours for an entire night).

The recipe is **gravity\_vis**.

The product (\*\_VIS) is an uncalibrated OIFITS file, containing the flux, the square visibilities, the complex visibilities and the closure phases for both the SC and the FT combiners.

The recipe produces an optional product, called P2VMRED, which contains many intermediate signals of the processing. It is very useful to assess the data quality. It is also possible to restart the reduction from the P2VMRED product, thus saving time for users who want to explore several values of recipe parameters.

### 3.5 From raw visibilities to calibrated visibilities

The second step is to calibrate the observation of science objects with the observation of calibration stars. This step could be performed “globally”, that is all the uncalibrated \*\_VIS oifits can be loaded first, then the pipeline searches for consistent sequences (same setup, same DIT, same wavelength table...), and then each sequence is calibrated independently with the calibration stars that could be find within it. This step is rather short (1min for an entire night).

The recipe is **gravity\_viscal**.

The product (\*\_VIS\_CALIBRATED) is a calibrated OIFITS file, supposedly ready for science.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	14 of 51

## 4 Instrument Data Description

### 4.1 RAW science data

The RAW frames when observing have the following DPR.TYPE:

<b>OBJECT,DUAL</b>	are observations of a nearby pair of object, one feeding the fringe-tracker (FT) and the other feeding the science combiner (SC). It can be of category SCI or CAL.
<b>OBJECT,SINGLE</b>	are observations of a single object, feeding both the fringe-tracker (FT) and the science combiner (SC). It can be of category SCI or CAL.
<b>SKY,SINGLE</b> <b>SKY,DUAL</b>	are observation of an empty patch of the sky near the object in order to measure the sky brightness.

The OBJECT,DUAL and OBJECT,SINGLE types have a category DPR.CATG=SCI when observing a science target, and a category DPR.CATG=CAL when observing a calibration star used to monitor the transfert function.

### 4.2 RAW calibration data

The RAW frames used to calibrate the instrument on a daily-basis have the following DPR.TYPE:

<b>DARK</b>	are observations with all shutters closed, in order to calibrate the detector dark illumination and the detector + dark illumination noise.
<b>FLAT</b>	are observation of the internal source with one shutter open, in order to calibrate the positions of the spectra on the detectors and the internal transmission of the instrument.
<b>P2VM</b>	are observations of the internal source with two shutters open, in order to calibrate the internal contrasts and phases of the instrument.
<b>WAVE</b> <b>WAVESC</b>	are observations of the internal source with all shutters open, in order to calibrate the wavelength table, and the internal closure phases. The WAVE data are recorded by scanning on both SC and FT, with FDDL in open loop and no fringe tracking (in order to calibrate the FT wavelengths). The WAVESC data are recorded by scanning only SC, and FDDL in close-loop and fringe tracking (to calibrate the SC wavelengths).

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	15 of 51

### 4.3 STATIC calibration

The STATIC calibration frames have the following DPR.TYPE:

<b>DISP_MODEL</b>	is the model of the optical dispersive index $n(\lambda)$ of the fibered differential delay lines (FDDL) of the instrument.
<b>DIAMETER_CAT</b>	is the catalog of stellar diameter, used to estimate the transfer function.
<b>EOP_PARAM</b>	is a list of Earth Orientation Parameters (EOP) and DUT1 versus time. These corrections are only needed for the most demanding astrometric measurements.

### 4.4 PRODUCT calibration data

The PRODUCT of the calibration by the recipes **gravity\_dark** and **gravity\_p2vm** are identified by the following PRO.CATG keyword:

<b>DARK</b>	contains images with the dark level and variance for the SC and FT detectors.
<b>BAD</b>	contains images with the identified bad pixels for the SC and the FT detectors.
<b>FLAT</b>	contains images of the profiles used to extract the SC spectra from the detector. There is one extracted spectrum per output of the detector and per polarisation if split (thus 24 or 48 spectra for each SC and FT combiners).
<b>WAVE</b>	contains tables with the effective wavelengths of every channel of every spectra extracted with the profile. These tables are necessary to re-align the different spectra (output of the detector) into a common wavelength grid.
<b>P2VM</b>	contains tables with the internal transmission, contrast and phase of every output of the detector versus wavelength. These form the so-called pixel-2-visibility matrix used to extract the interferometric visibility from the spectra.

### 4.5 PRODUCT science data

The PRODUCT of the science reduction by the recipe **gravity\_vis** are identified by the following PRO.CATG keyword:

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	16 of 51

**SINGLE\_SKY**  
**DUAL\_SKY**

contains the mean brightness of the sky and its variance. Contrarily to DARKs, these quantities possibly depend on the instrument mode because different optics are used.

**SINGLE\_SCI\_VIS**  
**SINGLE\_CAL\_VIS**  
**DUAL\_SCI\_VIS**  
**DUAL\_CAL\_VIS**

are OIFITS files [4] with the uncalibrated flux, square visibilities, complex visibilities and closure phases extracted from the raw observation of an object. SCI/CAL corresponds to a science target or a calibration star used to monitor the transfert function.

**SINGLE\_SCI\_P2VMRED**  
**SINGLE\_CAL\_P2VMRED**  
**DUAL\_SCI\_P2VMRED**  
**DUAL\_CAL\_P2VMRED**

are the raw data already processed through the P2VM algorithm. They contain flux per beam, and coherent flux per baseline, for each individual frame of the exposure. As such, they are intermediate products between the RAW data and the final, averaged, OIFITS. They also contains many internal signals of the processing. The file size is huge (>200Mb). It is meant to assess the overall data quality and tune the reduction parameters. It is not used for science. Its format is inspired by OIFITS, but it is not strictly compliant.

**SPECTRUM**  
**PREPROC**

contains the RAW data already cleanup from cosmetic and collapsed into one spectrum per combiner output. In SPECTRUM, the data are not yet re-interpolated into a common spectral wavelength grid, while this step is done in PREPROC. As such, they are debug-level intermediate products between the RAW data and the final, averaged, OIFITS.

The PRODUCT of the final calibration step by the recipe **gravity\_viscal** are identified by the following PRO.CATG keywords:

**SINGLE\_SCI\_VIS\_CALIBRATED**  
**DUAL\_SCI\_VIS\_CALIBRATED**

are the final OIFITS file of the reduction, science ready. They contains the interferometric observations calibrated from the transfer function.

**SINGLE\_CAL\_TF**  
**DUAL\_CAL\_TF**

are OIFITS files containing the transfer function value estimated by the corresponding observation of a calibration star. It is the observed visibility of a calibrator divided by its visibility estimated from its diameter.

**SINGLE\_SCI\_TF**  
**DUAL\_SCI\_TF**

are OIFITS files containing the estimated (interpolated) value of the transfer function at the time of the corresponding science target observation.



ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	17 of 51

## 5 Data Reduction

### 5.1 Graphical overview of the cascade




Figure 5.1: Data reduction cascade for observation in mode SINGLE. As of now, the cascade is the same for the mode DUAL.

### 5.2 Using Gasgano

*Gasgano*, provides a graphic interface for data browsing, classification and association, and offers several other utilities such as easy access to recipes documentation and preferred data display tools.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	18 of 51

*Gasgano* can be started from the system prompt in the following way:

```
gasgano &
```

Use the *Add/Remove Files* entry of the *File* menu to load data. The data are hierarchically organised as preferred by the user. More information about a single frame can be obtained by clicking on its name.

Frames can be selected from the main window for being processed by the appropriate recipe. Before launching the recipe, its configuration may be modified on the *Parameters* panel (on top). The window contents might be saved for later use by selecting the *Save Current Settings* entry from the *File* menu.

Please refer to the *Gasgano User's Manual* [7] for a more complete description of the *Gasgano* interface.

### 5.3 Using EsoRex

*EsoRex* is a command line utility for running pipeline recipes. It may be embedded by users into data reduction scripts for the automation of processing tasks. The user are let free to define manually the input SOF and the appropriate configuration parameters.

A SOF for *EsoRex* is a simple ASCII file listing the files and their *DO Category*. Exemple of SOF are given in the description of each recipe. Note that '#' is the comment character.

The basic syntax to use *EsoRex* is the following:

```
esorex [esorex_options] recipe_name [recipe_options] set_of_frames.sof
```

To get more information on how to customise *EsoRex* (see also [7]), or on a specific recipe, run the commands:

```
esorex -h
esorex -h recipe_name
esorex --man-page recipe_name
```

For more information on *EsoRex*, see <http://www.eso.org/cpl/esorex.html>.

### 5.4 Using run\_gravi\_reduce.py python script

The consortium has written a simple python script that classify the files in the current directory, associate them with calibrations, and run the corresponding recipes.

**Instrument calibration and data reduction** is triggered by the following script:

```
run_gravi_reduce.py [options]
```

The script performs the following steps:

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	19 of 51

1. Trigger the recipe **gravity\_dark** on all standalone DARK.
2. Trigger the recipe **gravity\_p2vm** on all sequence of 1DARK,4FLAT,6P2VM,1WAVE,1WAVESC.
3. Trigger the recipe **gravity\_vis** on all OBJECT,\*.

The script will search for (static) calibration in the current directory if any. It will also search in the upper directory `../common_calibration/`. The products are created in the `reduced/` subdirectory.

Before triggering a recipe, the script writes the corresponding SOF and the esorex command in the `reduced/` subdirectory. Thus one can manually reproduce a given reduction by executing:

```
./reduced/GRAVI.XXXX_esorex.sh
```

To get more information on syntax and options, run the commands:

```
run_gravi_reduce.py -h
```

**Transfer function and trending** can be then triggered with the following script:

```
cd reduced/
run_gravi_trend.py [options]
```

The script performs the following steps:

1. Trigger the recipe **gravity\_viscal** on all \*CAL\_VIS.
2. Trigger the recipe **gravity\_viscal** on all \*SCI\_VIS.
3. Produce trending plots to check the transfer function.

The products are created in the `calibrated/` and `trend/` subdirectories.

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	20 of 51

## 6 Known Problems

### 6.1 Spectral calibration

The spectral calibration between baselines is accurate to 1nm. It corresponds to is half a pixel (thus 1/4 of the spectral resolution element) in HR.

This uncertainty can generate biases in the closure phase, which amount to  $\approx 3\text{deg}$  when observing at a group-delay of 40  $\mu\text{m}$ .

The absolute spectral calibration is accurate to 2nm, which correspond to half a spectral resolution element in HR.

### 6.2 Uncertainties in products

The uncertainty on product contains the statistical noise only, computed by bootstrapping over the NDIT when possible. It does not contain the calibration uncertainty.

When the number of valid DIT within an exposure is lower than 5, the statistic to compute the final error bars also include additional MonteCarlo realisation of the *theoretical* photon and detector noise (to reach 5 samples). These uncertainties are thus less realistic.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	21 of 51

## A Pipeline Recipes Interfaces

### A.1 List of all recipes

We here list the role of each recipe. The input, output, options and QC parameters are detailed in each dedicated subsection.

gravity_badpix	Calibrate the badpixels from the detectors.
gravity_biasmask	*UNOFFERED* Determine which pixels can be used to measure the bias of SC detector.
gravity_dark	Calibrate the detector noise and background level.
gravity_disp	Calibrate the linearity and the dispersion of the differential delay lines.
gravity_eop	Download the last values of the Earth Orientation Parameters and DUT from IERS.
gravity_image	*UNOFFERED* Reconstruct an image from visibilities.
gravity_nab	*UNOFFERED* Calibrate the narrow angle baseline.
gravity_p2vm	Calibrate the instrument bad pixels, wavelength table, interferometric contrast and phase.
gravity_piezo	*UNOFFERED* Calibrate the response of the piezo actuators.
gravity_postprocess	Post-process the products, to fine-tune their content.
gravity_vis	Compute the visibilities from raw observation of OBJECT.
gravity_vis_from_p2vmred	Compute the visibilities from P2VMRED intermediate product.
gravity_viscal	Calibrate visibilities from the transfer function.
gravity_wavelamp	Measure the position of the Argon lines in the spectra.

### A.2 gravity\_dark

This recipe computes the DARK calibration for the SC and the FT detector. The SC detector is first debias using the biaspixels, before computing the dark mean and rms. For both detector, the mean dark level of each pixel and the stdev of each pixel are saved in the output product.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	22 of 51

## Input

DO.CATG	short description
DARK_RAW	raw dark, all shutters closed (DPR.TYPE=DARK)

## Output

PRO.CATG	short description
DARK	dark calibration (PRO.CATG=DARK)

## Parameters

Name	short description
–static-name –bias-method	Use static names for the products (for ESO). [FALSE] Methode to average the biaspixels when cleaning-up the SC detector (only apply to MED and LOW). Ideally the same value shall be used when reducing the DARK with gravity_dark and the OBJECT with gravity_vis. <MEDIAN   MEDIAN_PER_COLUMN> [MEDIAN]

## Quality control

QC in DARK	short description
PIXBIAS AVG	Mean of the pixels used to removed the detector bias. This value shall be add to MEDIANDARK SC to trend the detector dark illumination [adu].
PIXBIAS RMS	Std over the pixels used to remove the detector bias [adu].
MEDIANDARK SC	Median of the dark level in the Science Combiner detector [adu]
DARKRMS SC	Median of the dark rms (detector noise) in the Science detector [adu]
MEANDARK FT	Mean of the dark level in the FT detector [adu]
DARKRMS FT	Median of the dark rms (detector noise) in the Fringe-Tracker detector [adu]

## A.3 gravity\_p2vm

This recipe reduces the internal calibrations. The recipe is normally triggered on a SOF with RAW files only, but one can also input a SOF with some already processed calibration (e.g WAVE). The recipe is normally triggered with a complete SOF, but one can also input a SOF with DARK\_RAW only, or DARK\_RAW and FLAT\_RAW only.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	23 of 51

## Input

DO.CATG	short description
DARK_RAW	raw dark, all shutters closed (DPR.TYPE=DARK)
FLAT_RAW x4	raw flats, one shutter open (DPR.TYPE=FLAT)
P2VM_RAW x6	raw p2vms, two shutters open (DPR.TYPE=P2VM)
WAVE_RAW	raw wavelength calibration for FT (DPR.TYPE=WAVE)
WAVESC_RAW	raw wavelength calibration for SC (DPR.TYPE=WAVESC)

## Output

PRO.CATG	short description
DARK	dark calibration (PRO.CATG=DARK)
FLAT	flat calibration (PRO.CATG=FLAT)
BAD	badpixel calibration (PRO.CATG=BAD)
WAVE	wave calibration (PRO.CATG=WAVE)
P2VM	p2vm calibration (PRO.CATG=P2VM)

## Parameters

The parameters of this recipe are more debug-oriented rather than user-oriented. It is strongly advised to not modified them without reading the code first.

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]
–debug-file	Save additional debug file(s). [FALSE]
–preproc-file	Save the PREPROC intermediate product. [FALSE]
–bad-dark-threshold	the rms factor for dark bad pixel threshold. [10]
–force-badpix-to-zero	Force the badpixel to zero in profile. [TRUE]
–profile-width	width of profile in pixel. [6]
–phase-calibration	The relative phase of the P2VM are defined by NONE defines phiA(lbd) at zero for all baselines; CLOSURE defines phiA(lbd) at zero for baselines (01,02,03); DISP defines phiA(lbd) to have zero mean and minimum GD for baselines (01,02,03); FULL defines phiA(lbd) to have zero-GD for baselines (01,02,03). <NONE   CLOSURE   DISP   FULL> [CLOSURE]

## Quality control

QC in BAD	short description
BADPIX SC	Total number of bad pixels on the SC detector

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	24 of 51

BADPIX FT	Total number of bad pixels on the FT detector
BADPIX_DARK SC/FT	Pixels with weird mean level
BADPIX_RMS SC/FT	Pixels with weird noise level
BADPIX_FLAT SC/FT	Pixels non-responding to illumination
QC in FLAT	short description
PROFILE_CENTER SC1 MED	[pixel] position of the first spectra on SC detector
PROFILE_WIDTH SC1 MED	[pixel] width of the first spectra on SC detector
PROFILE_CENTER SC13 MED	[pixel] position of the 13d spectra on SC detector
PROFILE_WIDTH SC13 MED	[pixel] width of the 13d spectra on SC detector
MEANGAIN SC	Mean gain [ADU/e] for SC detector
MEANGAIN FT	Mean gain [ADU/e] for FT detector
QC in P2VM	short description
FLUX_SCi AVG	[e/DIT/chanel/output] flux in SC (mean of files)
FLUX_FTi AVG	[e/DIT/chanel/output] flux in FT (mean of files)
P2VM_COHERENCE_AVG_SC	Average instrumental contrast of SC
P2VM_COHERENCE_AVG_FT	Average instrumental contrast of FT
P2VM_COHERENCE_SCij	Average instrumental contrast of SC for pair ij
P2VM_COHERENCE_FTij	Average instrumental contrast of FT for pair ij
QC in WAVE	short description
REFWAVE1	Reference wavelength [m] for the below parameters
REFPOS1 SCi	Position [pix] of the REFWAVE1 in output SCi
REFPOS1 FTi	Position [pix] of the REFWAVE1 in output FTi
REFWAVE2	Reference wavelength [m] for the below parameters
REFPOS2 SCi	Position [pix] of the REFWAVE2 in output SCi
REFPOS2 FTi	Position [pix] of the REFWAVE2 in output FTi
WAVE_CORR	Model to convert the glass wavelength in vacuum wavelength
WAVE_CORR N0	Paramater of above model
WAVE_CORR N1	Paramater of above model
WAVE_CORR N2	Paramater of above model
MINWAVE SC/FT	Min wavelength [m] of SC/FT channels
MAXWAVE SC/FT	Max wavelength [m] of SC/FT channels
RMSWAVE SC/FT	Rms of residuals during polynomial wavelength fit

## A.4 gravity\_eop

This recipe download the latest version of the Earth Orientation Parameter and DUT from the IERS site. A web connection is required. The recipe doesn't need any input, but it creates the EOP\_PARAM product.

### Parameters

Name	short description
-eop_host	FTP Host to retrieve the EOP from. [ftp.iers.org]
-eop_urlpath	FTP URL path of the EOP file to retrieve. [/products/eop/rapid/standard/finals2000A.data]



<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	25 of 51

## Output

PRO.CATG	short description
EOP_PARAM	A file with the Earth Orientation Parameters and DUT versus time

### A.5 gravity\_vis

This recipe is associated to the observing template. It reduces the raw data acquired on calibrator or science targets and computes the uncalibrated visibilities, saved in an OIFITS file. If several OBJECT are provided, the recipe will reduce all of them and merge the resulting data into a single OIFITS. If several SKY\_RAW are provided, the recipe reduces the first OBJECT with the first SKY. Then each new OBJECT with the next SKY. When the number of sky is reached, the recipe loops back to first sky (so if the number of SKYs is larger than the number of OBJECTs, the last SKY won't be used). The recipe will reduce the data even if no SKY or no DARK is provided. However this will lead to wrong estimate of the visibility and square visibility of the object. If the DIAMETER\_CAT is not provided, the recipe will use the diameter provided in the header to compute the transfer function QC parameters. The tag in the DO category can be SINGLE/DUAL and CAL/SCI. They should reflect the mode (SINGLE or DUAL) and the DPR.CATG of the observation (SCIENCE or CALIB). The tag in the PRO.CATG category will be SINGLE/DUAL and CAL/SCI depending on the input tag.

## Input

DO.CATG	short description
FLAT	flat calibration (PRO.CATG=FLAT)
BAD	badpixel calibration (PRO.CATG=BAD)
WAVE	wave calibration (PRO.CATG=WAVE)
P2VM	p2vm calibration (PRO.CATG=P2VM)
DARK	dark calibration (PRO.CATG=DARK)
SINGLE_SCI_RAW	raw object (DPR.TYPE=OBJECT,SINGLE)
SINGLE_SKY_RAW	raw sky (DPR.TYPE=SKY,SINGLE)
DISP_MODEL (opt)	fiber dispersion model (PRO.CATG=DISP_MODEL)
DIAMETER_CAT (opt)	catalog of diameter (PRO.CATG=DIAMETER_CAT)

## Output

PRO.CATG	short description
SINGLE_SCI_VIS	OIFITS with uncalibrated visibilities
SINGLE_SKY (opt)	sky map
SINGLE_SCI_P2VMRED (opt)	intermediate product
SPECTRUM (opt)	intermediate product
PREPROC (opt)	intermediate product

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	26 of 51

## Parameters

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]
–bias-subtracted-file	Save the BIAS_SUBTRACTED intermediate product. [FALSE]
–spectrum-file	Save the SPECTRUM intermediate product. [FALSE]
–preproc-file	Save the PREPROC intermediate product. [FALSE]
–p2vmreduced-file	Save the P2VMRED intermediate product. [FALSE]
–astro-file	Save the ASTROREDUCED intermediate product. [FALSE]
–average-vis	Average the observation from the different input files (if any) in the output product, instead of simply appending them. [FALSE]
–bias-method	Methode to average the biaspixels when cleaning-up the SC detector (only apply to MED and LOW). Ideally the same value shall be used when reducing the DARK with gravity_dark and the OBJECT with gravity_vis. <MEDIAN   MEDIAN_PER_COLUMN> [MEDIAN]
–nsmooth-snr-ft	Number of sample to average coherently when computing the real-time SNR and GDELAY of the FT (shall corresponds to the atmospheric coherence time). The runing integration window is actually -nsmooth -> +nsmooth. [5]
–snr-min-ft	SNR threshold to accept FT frames (>0). It rizes the first bit (<<0) of column REJECTION_FLAG of FT. [3.0]
–state-min-ft	Minimum OPDC state to accept FT frames (>=0) It rizes the second bit (<<1) of column REJECTION_FLAG of FT. [1.0]
–tracking-min-sc	Minimum ratio of accepted FT frames to accept SC frames (0..1), that is, for each SC DIT, the fraction of the time the REJECTION_FLAG of the FT is not 0. It rizes the first bit (<<0) of column REJECTION_FLAG of SC. [0.8]
–vfactor-min-sc	vFactor threshold to accept SC frame (0..1). [0.1]
–debias-sc	Subtract the V2 bias from SC. [TRUE]
–debias-ft	Subtract the V2 bias from FT. [TRUE]
–nboot	Number of bootstrap to compute error (1..100). [20]
–vis-correction	Correction of SC visibility from the losses du to long-exposure. Either using the measured visibility losses with the FT (VFACTOR), or by forcing the SC visibilities to match those of the FT (FORCE). <VFACTOR   FORCE   NONE> [VFACTOR]
–flat-flux	Flat the OI_FLUX with instrument transmission recorded in the input P2VM calibration map. Thus flux is the spectrum recorded at the detector (FALSE); or the spectrum at the instrument entrance (TRUE). [FALSE]

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	27 of 51

-average-sky	Average the SKYs into a master SKY. If FALSE, the recipe loops over the SKY to reduce each OBJECT with a different SKY. [FALSE]
--------------	---

## Quality control

QC in VIS	short description
CHECK FLAGS	Number of message about data integrity
CHECK MSGi	Message about data integrity
TRANS PROFILE SC	[e/e] numerical transmission at profile extraction
TRANS INTERP SC	[e/e] numerical transmission at interpolation
TRANS P2VM SC	[e/e] numerical transmission at P2VM extraction
PHASE_FTij RMS	[rad] phase residuals when tracking
TRACKING_RATIO_FTij	[%] ratio of tracking time
ACCEPTED_RATIO_FTij	[%] fraction of accepted FT frame in the final averaging
ACEEPTED_RATIO_SCij	[%] fraction of accepted SC frame in the final averaging
TAU0 OPDCij	[s] tau0 for variance of 1 rad <sup>2</sup> , computed from the fringe-tracking command to the piezo actuator.
VFACTORij_P1 AVG	mean v-factor for polar 1
VISAMP_SCij_P1 AVG	[deg] mean VISAMP for Science Combiner for polar 1
...	see all parameters
GD_SCij_P1 AVG	[m] Mean Group-Delay on Science Combiner
FLUX_FTi_P1 AVG	[e/total_int_time] mean flux over channels for Fringe Tracker
FLUXRATE_FTi_P1 SUM	[e/s] sum over channels for Fringe Tracker

If the observation is of type CALIB, and if the DIAMETER\_CAT was provided, the recipe also computes the following QC:

QC in VIS	short description
TF TRANS_FTij	Total transmission of FT
TF TRANS_SCij	Total transmission of SC
TF VISMOD_SCij RELERR	TF relative error from diameter error
TF VISAMP_SCij_P1 MED	TF median over channels
TF VIS2_SCij_P1 MED	TF median over channels

## A.6 gravity\_vis\_from\_p2vmred

This recipe average the real-time data from input P2VMRED files into a VIS product. It allows to run the reduction with different parameters (for instance for SNR thresholding) without having to re-reduce the files from scratch. Typically the reduction is 4x faster when started from this intermediate product. The tag in the DO category can be SINGLE/DUAL and CAL/SCI. They should reflect the mode (SINGLE or DUAL) and the DPR.CATG of the observation (SCIENCE or CALIB). The tag in the PRO.CATG category will be SINGLE/DUAL and CAL/SCI depending on the input tag.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	28 of 51

## Input

DO.CATG	short description
SINGLE_SCI_P2VMRED	Input intermediate product

## Output

PRO.CATG	short description
SINGLE_SCI_VIS	OIFITS with uncalibrated visibilities

## Parameters

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]
–average-vis	Average the observation from the different input files (if any) in the output product, instead of simply appending them. [FALSE]
–nsmooth-snr-ft	Number of sample to average coherently when computing the real-time SNR and GDELAY of the FT (shall corresponds to the atmospheric coherence time). The runing integration window is actually -nsmooth -> +nsmooth. [5]
–snr-min-ft	SNR threshold to accept FT frames (>0). It rizes the first bit (<<0) of column REJECTION_FLAG of FT. [3.0]
–state-min-ft	Minimum OPDC state to accept FT frames (>=0) It rizes the second bit (<<1) of column REJECTION_FLAG of FT. [1.0]
–tracking-min-sc	Minimum ratio of accepted FT frames to accept SC frames (0..1), that is, for each SC DIT, the fraction of the time the REJECTION_FLAG of the FT is not 0. It rizes the first bit (<<0) of column REJECTION_FLAG of SC. [0.8]
–vfactor-min-sc	vFactor threshold to accept SC frame (0..1). [0.1]
–debias-sc	Subtract the V2 bias from SC. [TRUE]
–debias-ft	Subtract the V2 bias from FT. [TRUE]
–nboot	Number of bootstrap to compute error (1..100). [20]
–vis-correction	Correction of SC visibility from the losses du to long-exposure. Either using the measured visibility losses with the FT (VFACTOR), or by forcing the SC visibilities to match those of the FT (FORCE). <VFACTOR   FORCE   NONE> [VFACTOR]
–use-existing-rejection	Use existing rejection flags (ignore related options). [FALSE]

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	29 of 51

## A.7 gravity\_viscal

This recipe calibrates the visibilities acquired on science target using visibilities acquired on calibrator target. If the DIAMETER\_CAT is not provided, the recipe will use the diameter provided in the header to compute the transfer function QC parameters. The corresponding keywords are INS.SOBJ.DIAMETER and FT.ROBJ.DIAMETER. The OI\_FLUX data are not yet calibrated. The tag in the DO category can be SINGLE/DUAL and CAL/SCI. They should reflect the mode (SINGLE or DUAL) and the DPR.CATG of the observation (SCIENCE or CALIB). The tag in the PRO.CATG category will be SINGLE/DUAL and CAL/SCI depending on the input tag.

### Input

DO.CATG	short description
SINGLE_SCI_VIS (>=1)	visibilities on sciences
SINGLE_CAL_VIS (>=1)	visibilities on calibrators
DIAMETER_CAT (opt)	catalog of diameter

### Output

PRO.CATG	short description
SINGLE_SCI_VIS_CALIBRATED	calibrated science visibilities
SINGLE_CAL_TF	Transfer Function (TF) estimated on calibrators
SINGLE_SCI_TF	TF interpolated at the time of sciences

### Parameters

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]
–delta-time-calib	Delta time to interpolate the TF [s]. [3.6e+03]
–force-calib	Force the calibration, don't check setup. [FALSE]

## A.8 gravity\_postprocess

The recipe allows to manipulate the product of the GRAVITY pipeline, mostly the VIS. It permits to merge several files together into a single VIS file with all observations; to average the observations of one or several VIS file to increase the SNR; to remove some data (FT, SC); and to resample the SC observation with spectral binning.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	30 of 51

## Input

The list of input file can be P2VMRED, VIS, VIS\_CALIBRATED (or even RAW for some parameters). However they should all be compatible in term of setup and observed objet !! Note that the recipe performs only little checks of the input file content and structure. Thus the user shall ensure the input files are conformable (same polarisation and spectral mode for instance)

## Parameters

Name	short description
–average-vis	Average the observation from the different input files (if any) in the output product, instead of simply appending them. [FALSE]
–force-merge	Force merging even if inconsistent data. [FALSE]
–remove-ft	Remove FT extensions. [FALSE]
–remove-sc	Remove SC extensions. [FALSE]
–remove-opdc	Remove OPDC extensions. [FALSE]
–remove-met	Remove METROLOGY related extensions. [FALSE]
–nbin-lambda-sc	Bin SC extensions in spectral dimension. [0]

## Output

PRO.CATG	short description
POSTPROCESSED	The result of all input file.

## A.9 gravity\_wavelamp

This recipe is associated to the template gravity\_wavelamp. It reduce the raw argon file (WAVELAMP) and process it.

## Input

DO.CATG	short description
FLAT	flat calibration (PRO.CATG=FLAT)
BAD	badpixel calibration (PRO.CATG=BAD)
WAVE	wave calibration (PRO.CATG=WAVE)
P2VM	p2vm calibration (PRO.CATG=P2VM)
WAVELAMP_RAW	long exposure of Argon lamp
DARK_RAW	dark of Argon exposure

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	31 of 51

## Output

PRO.CATG	short description
WAVELAMP	spectrum of Argon, with position of lines

## Parameters

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]

### A.10 gravity\_disp

This recipe is associated to the template GRAVI\_all\_disp. It measure the phases obtained on the internal source at the position of the Argon lines and various stretch of the FDDL. It deduces the linearity model and the dispersion model of the differential delay lines. These models are stored as polynomials versus wavelength.

## Input

DO.CATG	short description
FLAT	flat calibration (PRO.CATG=FLAT)
BAD	badpixel calibration (PRO.CATG=BAD)
WAVE	wave calibration (PRO.CATG=WAVE)
P2VM	p2vm calibration (PRO.CATG=P2VM)
DARK	dark calibration
DISP_RAW >10	raw dispersion

## Output

PRO.CATG	short description
DISP_MODEL	dispersion model of FDDL

## Parameters

Name	short description
–static-name	Use static names for the products (for ESO). [FALSE]
–preproc-file	Save the PREPROC intermediate product. [FALSE]
–p2vmreduced-file	Save the P2VMRED intermediate product. [FALSE]

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	32 of 51

–nsmooth-snr-ft	Number of sample to average coherently when computing the real-time SNR and GDELAY of the FT (shall corresponds to the atmospheric coherence time). The runing integration window is actually -nsmooth -> +nsmooth. [5]
–snr-min-ft	SNR threshold to accept FT frames (>0). It rizes the first bit («0) of column REJECTION_FLAG of FT. [30.0]
–state-min-ft	Minimum OPDC state to accept FT frames (>=0) It rizes the second bit («1) of column REJECTION_FLAG of FT. [1.0]
–tracking-min-sc	Minimum ratio of accepted FT frames to accept SC frames (0..1), that is, for each SC DIT, the fraction of the time the REJECTION_FLAG of the FT is not 0. It rizes the first bit («0) of column REJECTION_FLAG of SC. [0.8]
–vfactor-min-sc	vFactor threshold to accept SC frame (0..1). [0.8]
–debias-sc	Subtract the V2 bias from SC. [TRUE]
–debias-ft	Subtract the V2 bias from FT. [TRUE]
–nboot	Number of bootstrap to compute error (1..100). [1]
–vis-correction	Correction of SC visibility from the losses du to long-exposure. Either using the measured visibility losses with the FT (VFACTOR), or by forcing the SC visibilities to match those of the FT (FORCE). <VFACTOR   FORCE   NONE> [NONE]



<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	33 of 51

## B Re-creating the Static Calibration

### B.1 Dispersion model

The **DISP\_MODEL** static calibration can be recomputed with the recipes **gravity\_wavelamp** and **gravity\_disp**. The principle is to accurately measure the interferometric phases obtained for various stretching of the instrument fibers at the wavelengths of known Argon lines. The following dedicated RAW data are required:

**WAVELAMP** is a spectra of the internal argon lamp.

**DISP** are interferometric observation of the internal source with various stretch of the fibers.

### B.2 Earth Orientation Parameters

The **EOP\_PARAM** static calibration can be recomputed with the recipe **gravity\_retrieve\_eop**, which shall query the IERS webpage to obtain the best estimate of the past and futur EOP.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	34 of 51

## C Detailed description of the data content

### C.1 Recommended tools to browse data

RAW files can be conveniently open with the last version of *QFitsView* [2] from Thomas Ott.

OIFITS product files can be conveniently open with the last version of *OIFits Explorer* [1] from the JMMC.

The consortium has built a python script to visualise most of the GRAVITY data, *run\_gravi\_visual.py*.

### C.2 Table structure common to all data

The INSNAME header keyword gives the combiner to which a table refers to, and thus allows to cross-reference with other tables. The EXTVER keyword code the same information on an integer to allow fast search in the tables (see python FITS class for instance). They can take the values: GRAVITY\_FT' (10), GRAVITY\_FT\_P1' (11), GRAVITY\_FT\_P2' (12), GRAVITY\_SC' (20), GRAVITY\_SC\_P1' (21), GRAVITY\_SC\_P2' (22).

The IMAGING\_DETECTOR\_SC and IMAGING\_DETECTOR\_FT tables store the detector configurations following the OIFITS standard [4].

The IMAGING\_DATA\_SC and IMAGING\_DATA\_FT extensions store the detector data. The SC data are stored as image list, while the FT data are stored as tables.

The OI\_WAVELENGTH tables store the wavelength table following the OIFITS standard. They shall be associate to the SC or FT using the INSNAME or EXTVER keywords.

OIFITS tables storing quantities per-beam (OI\_FLUX) have a total of NDIT x 4 rows (or NEXP x 4 rows for final product). In these tables, the four beams are always ordered following 12341234...

OIFITS tables storing quantities per-baseline (OI\_VIS, OI\_VIS2) have a total of NDIT x 6 rows (or NEXP x 6 rows for final product). In these tables, the six baselines are always ordered following 123456123456... The baselines 1 to 6 are always the pair of the beams 1-2,1-3,1-4,2-3,2-4,3-4.

### C.3 P2VM product

The P2VM\_SC and P2VM\_FT tables are inspired from the OIFITS format, and contains the following columns:

**REGNAME** : Detector region name, to match the IMAGING\_DATA table.

**TRANSMISSION** : For each region (= output of the combiner), a ntel × nwave image with the transmission of each input beam in this region. Since the combination scheme is pairwise, normaly only 2 rows of this image shall be non-zero.

**COHERENCE** : For each region (= output of the combiner), a nbase × nwave image with the instrumental visibility of each pair of input beam in this region. Since the combination scheme is pairwise, normaly only one single rows of this image shall be non-zero.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	35 of 51

**PHASE** : For each region (= output of the combiner), a  $n_{\text{base}} \times n_{\text{wave}}$  image with the instrumental phase of each pair of input beam in this region. Since the combination scheme is pairwise, normally only one single rows of this image shall be non-zero.

#### **C\_MATRIX :**

The OI\_FLUX tables store the internal instrument transmission, measured on the internal light, following the OIFITS standard, in arbitrary units. These tables are used to latter calibrate the measured flux in the object from the spectral shape of the internal transmission.

### **C.4 \*\_VIS and \*\_TF products**

The files with PRO.CATG=\*\_VIS and \*\_TF follow the OIFITS standard, version 1. All information can be found in [4].

The OIFITS tables contain one row per baseline and per corresponding RAW exposure, that is all the frames from the individual exposure are averaged together.

The OIFITS tables shall be associate to the SC or FT using the INSNAME (and/or EXTVER) keywords.

### **C.5 SPECTRUM, PREPROC products**

These are intermediate products used to debug the pipeline.

Both files contain the extracted spectrum of each output of the combiners. In SPECTRUM, the spectrum are in pixel space, thus on different wavelength grids. In PREPROC, the spectrum have been re-interpolated into a common wavelength grid.

#### **Columns in the SPECTRUM\_DATA tables**

**TIME** [us] : time of the frame, in [us], from the PRC.ACQ.START time from header (RMN recording start).

**DATAi** [e] : The spectrum of the flux from output i of the combiner.

**DATAERRi** [e] : The spectrum of the theoretical error on the flux from output i of the combiner, including detector and photonic variances.

### **C.6 \*\_P2VMRED products**

The files with PRO.CATG=\*\_P2VMRED are inspired from the OIFITS format [4], but are non-standard for the TIME colums. They also include many additional columns to store intermediate signal.

Beware that the OIFITS tables contain one row per frame: the number of row can be huge for the FT tables.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	36 of 51

### Columns in the OI\_VIS table of the SC

**TARGET\_ID** : id listed in OI\_TARGET

**TIME** [us] : time of the frame, in [us], from the PRC.ACQ.START time from header (RMN recording start).

**MJD** [day] :

**INT\_TIME** [s] : integration time of this frame

**VISDATA** [e,e] : complex coherent flux of SC in this frame

**VISERR** [e,e] : theoretical complex error on the coherent flux

**UCOORD** [m] : uv-plane of this SC frame

**VCOORD** [m] : uv-plane of this SC frame

**STA\_INDEX** : station index in the OI\_ARRAY

**PHASE\_MET\_FC** [rad] : unwrap FT-SC phase as computed by the DRS algorithm

**PHASE\_MET\_TEL** [rad] : unwrap FT-SC phase as computed by the DRS algorithm, mean of 4 diodes

**OPD\_MET\_FC** [m] : unwrap SC-FT delay as computed by the TAC algorithm

**OPD\_MET\_TEL** [m] : unwrap SC-FT delay as computed by the TAC algorithm, 4 diodes

**OPD\_DISP** [m] : spectra of OPD introduced by fibers, including dispersion.

**VISDATA\_FT** [e,e] : <VISDATA> spectra of FT (integrated in this SC frame)

**VISPOWER\_FT** [e\*\*2] : <|VISDATA|\*\*2> spectra of FT (integrated in this SC frame)

**VISVAR\_FT** [e\*\*2] : <|VISERR|\*\*2> spectra of FT (integrated in this SC frame)

**GDELAY\_FT** [m] : real-time GD computed from VISDATA\_FT

**PHASE\_REF** [rad] : reference phase, actually -1\*argVISDATA\_FT, re-interpolated in the SC wavelength.

**PHASE\_REF\_COEFF** [rad] : polynomial coefficients fit to argVISDATA\_FT and used to extrapolate to the SC wavelengths, in units of  $(\lambda - \lambda_{mean})/(\lambda_{max} - \lambda_{min})$

**FIRST\_FT** : first FT frame in this SC frame

**LAST\_FT** : last FT frame in this SC frame

**FIRST\_MET** : first MET frame in this SC frame

**LAST\_MET** : last MET frame in this SC frame

**GDELAY** [m] : real-time GD computed from VISDATA

**SNR** : real-time SNR

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	37 of 51

**GDELAY\_BOOT** [m] : best GD estimate, accounting closing triangles

**SNR\_BOOT** : best SNR estimate, accounting closing triangles

**V\_FACTOR\_FT** : measured visibility loss on the FT

**V\_FACTOR** : predicted visibility loss of this SC frame (re-interpolation of V\_FACTOR\_FT on the SC wavelengths)

**P\_FACTOR** : predicted visibility loss of this SC frame due to photometric flickering (based on the real-time photometry of the FT)

**F1F2** Estimate of geometric flux of this frame.

**FRINGEDET\_RATIO** : fraction of FT frame accepted in this SC frame

**REJECTION\_FLAG** : this frame is accepted/rejected

**E\_U, E\_V, E\_W** : Local celestial {u,v,w} (East, North, Toward observer = OIFITS standard) expressed in local terrestrial (East, North, Up) at Paranal reference. It is usefull to recompute the projected baseline from physical baseline.

**E\_Az** : Vectorial product between E\_W and the Zenith direction, expressed in the local terrestrial (East, North, Up) at Paranal reference. Sitting on telescope, looking at target, E\_Az points toward left in the horizontal plane.

**E\_Zd** : Vectorial product between E\_W and E\_Az, expressed in the local terrestrial (East, North, Up) at Paranal reference. Sitting on telescope, looking at target, E\_Az points toward Nadir in the plane perpendicular to pointing direction.

#### Columns in the OI\_VIS table of the FT

**TARGET\_PHASE** [rad] : target phase of the loop, including the Sylvester modulation

**STATE** : baseline tracking state as reported by OPDC

**GDELAY** [m] : real-time GD computed from VISDATA

**GDELAY\_BOOT** [m] : best GD estimate, accounting closing triangles

**SNR** : real-time SNR

**SNR\_BOOT** : best SNR estimate, accounting closing triangles

**REJECTION\_FLAG** : this frame is accepted/rejected

**F1F2** Estimate of geometric flux of this frame.

**PHASE\_REF** [rad] : self-reference phase.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	38 of 51

### Columns in the OI\_FLUX table of the SC

**TARGET\_ID** : id listed in OI\_TARGET

**TIME** [us] : time of the frame, in [us], from the PRC.ACQ.START time from header (RMN recording start).

**MJD** [day] :

**INT\_TIME** [s] : integration time of this frame

**FLUX** [e] : flux

**FLUXERR** [e] : theoretical error on flux

**STA\_INDEX** : station index in the OI\_ARRAY

**TOTALFLUX\_SC** [e] : total flux of SC in this SC frame (integrated over spectrum)

**TOTALFLUX\_FT** [e] : total flux of FT in this SC frame (integrated over spectrum)

**OPD\_MET\_FC** [m] : unwrap SC-FT delay as computed by the TAC algorithm

**OPD\_MET\_TEL** [m] : unwrap SC-FT delay as computed by the TAC algorithm, 4 diodes.

**FT\_POS** [V] : mean FT FDDL strain gauge voltage during this frame

**SC\_POS** [V] : mean SC FDDL strain gauge voltage during this frame

**FDDL** [m] : The mean of SC\_FDDL and FT\_FDDL, where these value are the FDDL strain gauge measurements, corrected from non-linearity and converted in [m].

**FIRST\_FT** : first FT frame in this SC frame

**LAST\_FT** : last FT frame in this SC frame

**FIRST\_MET** : first MET frame in this SC frame

**LAST\_MET** : last MET frame in this SC frame

**FIRST\_FDDL** : first FDDL frame in this SC frame

**LAST\_FDDL** : last FDDL frame in this SC frame

### Columns in the OI\_FLUX table of the FT

**STATE** : telescope state as reported by OPDC

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	39 of 51

### Columns in the OI\_VIS\_MET table

The table has the same structure as the OI\_FLUX table, that is one row per beam (12341234124...), and thus nsample x 4 rows.

**PHASE\_FC** [rad] : phases at combiner, unwrap by pipeline algorithm (FT-SC)

**PHASE\_TEL** [rad] : 4 diodes phases at telescope, unwrap by pipeline algorithm (FT-SC)

**OPD\_FC** [m] : OPD at telescope, unwrap by TAC algorithm (SC-FT)

**OPD\_TEL** [m] : 4 diodes OPD at telescope, unwrap by TAC algorithm (SC-FT)

**FLAG\_FC, FLAG\_TEL** : flags computed by TAC algorithm

**VAMP\_FC\_FT** , **VAMP\_FC\_SC**, **VAMP\_TEL\_FT**, **VAMP\_TEL\_SC**: Volt amplitude

### Columns in the OPDC table

**TIME** [us] : time of the frame, in [us], from the PRC.ACQ.START time from header (RMN recording start).

**PIEZO\_DL\_OFFSET**

**VLTI\_DL\_OFFSET**

**STATE** : global fringe tracking state

**BASELINE\_STATE** : fringe tracking state per telescope and baseline (scrambled)

**STEPS** : target phase modulation per baseline (scrambled), in units of pi/8

## C.7 ASTRORED product

This is a light version of the P2VMRED file. It is meant to develop the astrometric mode of GRAVITY.

## C.8 DISP\_MODEL

The data are stored in the table DISP\_MODEL. There are one line per GRAVITY beam. The columns are:

**N\_MEAN** : Mean optical index of the SC and FT fibers. The vector contains the  $a_i$  coefficients such that the index at wavelength  $\lambda$  is given by  $n(\lambda)/n(\lambda_0) = \sum_i a_i (\lambda - \lambda_0)^i / \lambda_0^i$

**N\_DIFF** : Differential optical index between the SC and the FT fibers. The vector contains the  $a_i$  coefficients such that the differential index at wavelength  $\lambda$  is given by  $\Delta n(\lambda)/n(\lambda_0) = \sum_i a_i (\lambda - \lambda_0)^i / \lambda_0^i$

**LIN\_FDDL\_FT** : Linearity coefficients to convert the FDDL signal in [V] to stretching length in [m]. The vector contains the  $a_i$  coefficients such that  $L(V) = \sum_i a_i V^i$ .

**LIN\_FDDL\_SC** : Same for the SC.

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	40 of 51

## D Algorithms

In the following,  $f$  is for the time (individual DIT, typically [0..30] for SC and [0..300000] for FT),  $ij$  are the pixels on the detector ( $i$  being the spatial direction, and  $j$  being the spectral direction),  $l$  is for the spectral channel,  $o$  is the combiner output (so called region in the code, [0..23] or [0..47] if polars are split),  $b$  is the baseline [0..5],  $t$  is the telescope/beam [0..4].

### D.1 Bias clean-up

The SC frames are clean-up using the bias pixels available. In LOW and MED spectral mode, the bias per frame is estimated as the mean over all the pixels of the *bias lines* interleaved between each region.

In HIGH spectral mode, the bias per frame is estimated as the mean over all pixels of the *bias columns* at the edge of the detector.

In both case, there is thus a single, scalar bias value for each frame.

### D.2 Spectrum extraction

The implemented spectrum extraction  $Y_{foj}$  from the 2D image  $X_{fij}$  is based on a profile image  $p_{oij}$ .

#### Profile definition

In LOW and MED spectral modes, the profile is directly the one observed with the sequences of FLAT files. It resembles a Gaussian function with FWHM of 1.5 pixel. To ensure the overall flux is conserved in the extraction, we apply the following normalization which assumes the shape of the data is perfectly matched by the profile itself:

$$p_{oij} = p_{oij} \cdot \frac{\sum_i p_{oij}}{\sum_i p_{oij}^2} \quad (1)$$

In HIGH mode, the profile is a box-card of 6 pixels around the center of the best-fit Gaussian on the observed profile in the FLAT files. To ensure flux conservation, this box-card is either 0 (outside) or 1 (inside).

#### Bad pixels in profile

The bad pixels are then forced to zero in the profile. Consequently, a profile with bad-pixel will lead to a reduced amount of detected flux. This effect is calibrated by the P2VM algorithm because the P2VM coefficient and the data are affected by the same amount of flux loss. A worst, for some spectral channel, the spectra of one output (e.g A) can be forced zero if all the pixels are bad. The P2VM then relies on the remaining BCD outputs only.

#### Extracted spectrum

$$Y_{foj} = g \sum_i (X_{fij} - S_{ij}) p_{oij} \quad (2)$$



<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	41 of 51

where  $S_{i,j}$  is the mean image measured on the SKY. The sum  $\sum$  is obviously performed in the compressed spatial direction.

Finally,  $g$  is the conversion gain from [ADU] to [e].

### Variance of the extracted spectrum

Introducing the photon and detector noise, the variance of the extracted spectrum is:

$$\text{var}(Y)_{foj} = g \sum_i (X_{fij} - D_{ij}) p_{oij}^2 + g^2 \sum_i \sigma_{ij}^2 p_{oij}^2 \quad (3)$$

where  $D_{i,j}$  is the mean image measured on the DARK, and  $\sigma_{ij}^2$  is the variance image measured on the DARK.

## D.3 Re-interpolation to a common wavelength

### Modified target wavelength for SC

For the output (region)  $o$ , let's call  $j_{ol}$  the sample just below the target wavelength  $\lambda_l$ , and  $j_{ol} + 1$  the sample just above. Of course  $j$  depends on the region because their wavelength tables are different.

For the SC, this target wavelength is slightly modified, in a different way, for each region:

$$\lambda'_{ol} = \lambda_{o,j_{ol}} + \frac{(\lambda_l - \lambda_{o,j_{ol}})(\lambda_{o,j_{ol}+1} - \lambda_{o,j_{ol}})}{(\lambda_l - \lambda_{o,j_{ol}}) + (\lambda_{o,j_{ol}+1} - \lambda_l) \cdot \frac{F_{o,j_{ol}+1}}{F_{o,j_{ol}}}} \quad (4)$$

where  $F_{oj}$  is the flat measured on the internal light, extracted the same way as the data. This modification ensures that we later interpolate to a common *effective wavelength* for all regions. It also ensures that spectral channels whose interpolation implies a bad-pixel (forced to zero) are all set to zero.

### Modified target wavelength for FT

For the FT, we don't modify the target wavelength:

$$\lambda'_{ol} = \lambda_l \quad (5)$$

### Interpolation of flux and variance

The following coefficient  $a_{ol}$

$$a_{ol} = \frac{\lambda_{o,j+1} - \lambda'_{ol}}{\lambda_{o,j+1} - \lambda_{o,j}} \quad (6)$$

allows to linearly interpolate the fluxes:

$$Y_{fol} = a_{ol} Y_{fo,j_{ol}} + (1 - a_{ol}) Y_{fo,j_{ol}+1} \quad (7)$$

and the variances:

$$\text{var}(Y)_{fol} = a_{ol}^2 \text{var}(Y)_{fo,j_{ol}} + (1 - a_{ol})^2 \text{var}(Y)_{fo,j_{ol}+1} \quad (8)$$

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	42 of 51

#### D.4 Extraction of the coherent fluxes and telescope fluxes via P2VM

The fluxes of each telescope  $F_{ftl}$  and the complexe coherent flux of each base  $R_{fbl} + iI_{fbl}$  are extracted from a matricial analysis of the profiles, based on the P2VM calibration

$$(F_{ftl}, R_{fbl}, I_{fbl}) = P2VM_{b/tl}^o \times Y_{fol} \quad (9)$$

The variances are propagated assuming no correlation between the input  $Y_{fol}$ .

$$(\text{var}(F)_{ftl}, \text{var}(R)_{fbl}, \text{var}(I)_{fbl}) = (P2VM_{b/tl}^o)^2 \times \text{var}(Y)_{fol} \quad (10)$$

#### D.5 Computation of SNR

##### Individual SNR

The SNR of each baseline and each frame of the the FT is computed using a running mean of the complex coherent flux over 10 samples. The coherent flux of the two polarisations, if any, are also averaged together after having recentered them to a common mean phase.

$$SNR_{fb} = \frac{(\sum_{f_r, l} R_{f_r, bl})^2 + (\sum_{f_r, l} I_{f_r, bl})^2}{\sum_{f_r, l} \text{var}(R)_{f_r, bl} + \sum_{f_r, l} \text{var}(I)_{f_r, bl}} \quad (11)$$

where  $f_r$  is in the interval  $\in \{f - 5, f + 5\}$

##### Bootstraped SNR

Then a “bootstraped” SNR is computed for each baseline and each frame, as the maximum between the SNR of this baseline and all closing triangles. For instance for the baseline  $b = b_{ij}$  among the beams  $i, j, k, l$ :

$$SNRB_{fb_{ij}} = \max\{ SNR_{fb_{ij}}, \min\{SNR_{fb_{ik}}, SNR_{fb_{kj}}\}, \min\{SNR_{fb_{il}}, SNR_{fb_{lj}}\} \} \quad (12)$$

The quantities SNR and SNRB are also computed for the SC, although they are not used in the processing.

#### D.6 Computing the vFactor

The purpose of the vFactor is to estimate the visibility loss of each individual SC frame from an analysis of the FT real-time data.

A white-light vFactor is first computed for each SC frame  $f$  with the FT data:

$$v_{fb} = \frac{(\sum_{f_{ft}} R_{f_{ft}bl})^2 + (\sum_{f_{ft}} I_{f_{ft}bl})^2 - \sum_{f_{ft}l} \text{var}(R)_{f_{ft}bl} - \sum_{f_{ft}l} \text{var}(I)_{f_{ft}bl}}{\sum_{f_{ft}} (\sum_l R_{f_{ft}bl})^2 + \sum_{f_{ft}} (\sum_l I_{f_{ft}bl})^2 - \sum_{f_{ft}l} \text{var}(R)_{f_{ft}bl} - \sum_{f_{ft}l} \text{var}(I)_{f_{ft}bl}} \times \frac{1}{n_{f_{ft}}} \quad (13)$$

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	43 of 51

where here the sum  $l$  is over the 5 spectral channels of the FT and the sum  $f_{\tau}$  is over the FT frames *inside* the corresponding SC frame. This white-light vFactor at  $\lambda_0$  (the central wavelength of the FT) is then extrapolated to the SC channels with:

$$\tilde{v}_{fbl} = \exp\left(-\ln(v_{fb}) \frac{\lambda_0^2}{\lambda_l^2}\right) \quad (14)$$

## D.7 Frame rejection

A FT frame is rejected if any of the conditions are met:

- its bootstrapped  $SNRB$  is below the threshold.
- the OPDC state of this baseline is below the threshold.

A SC frame is rejected if any of the conditions are met:

- the fraction of accepted FT frame during this SC frame is below the threshold.
- its computed vFactor is below the threshold.

The frame selection is done on a per-baseline basis. That is the baselines have a different selection map, and thus will have a different effective time after the averaging process.

## D.8 Phase referencing

### Referencing the SC phases with the FT

The phase reference of a SC frame is the mean phase of the FT during this SC frame:

$$\Upsilon_{fbl} = \arctan\left(\sum_{f_{\tau}} I_{f_{\tau}bl}, \sum_{f_{\tau}} R_{f_{\tau}bl}\right) \quad (15)$$

where the sum over  $f_{\tau}$  is over the FT frames *inside* the corresponding SC frame.

Then this  $\Upsilon$ , which have 5 spectral channels only, is projected into the wavelengths of the SC with polynomial fit of order 2 (after properly unwrapping the phase).

### Self-referencing the FT phase

The phase reference of a FT frames is the running mean phase of the FT itself over few samples:

$$\Upsilon_{fbl} = \arctan\left(\sum_{f_r} I_{f_rbl}, \sum_{f_r} R_{f_rbl}\right) \quad (16)$$

where  $f_r$  is in the interval  $\in \{f - 3, r + 3\}$ , excluding  $f_r = f$  to avoid biases.

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	44 of 51

## D.9 Geometric flux

The geometric flux of a baseline is the product of the photometric fluxes of the two beams of this baseline. It is necessary to then normalise the coherent fluxes into visibilities.

### For the FT

The FT real-time photometry is affected by large photometric fluctuations, which make it going near zero (flux loss event) or even go negative because of noise... To tackle this, we first compute a broad-band, temporally smoothed version of the real-time flux of each beam:

$$F'_{ft} = \sum_{f_r l} F_{f_r t l} \quad (17)$$

where  $f_r$  is in the interval  $\in \{f - 5, f + 5\}$  (time smoothing).

$$FF_{fbl} = \left( \sum_f F_{f t_1 l} \right) F'_{f t_1} \times \left( \sum_f F_{f t_2 l} \right) F'_{f t_2} \quad (18)$$

### For the SC

For the science, the geometric flux is simply computed as:

$$FF_{fbl} = F_{f t_1 l} \times F_{f t_2 l} \quad (19)$$

## D.10 Averaged flux estimator

All frames are averaged.

$$\widetilde{\text{flux}}_{tl} = \sum_f F_{f t l} \quad (20)$$

Hence the final flux is the *sum* of all electro-event collected during the entire exposure.

## D.11 Averaged complex visibility estimator

For a given baseline, the averaged is performed only with accepted frames of this baseline.

The complex coherent flux is first rotated with the computed reference phase:

$$R'_{fbl} = \cos(\Upsilon_{fbl}) R_{fbl} - \sin(\Upsilon_{fbl}) I_{fbl} \quad (21)$$

$$I'_{fbl} = \sin(\Upsilon_{fbl}) R_{fbl} + \cos(\Upsilon_{fbl}) I_{fbl} \quad (22)$$

The visibility of each frames are averaged together accounting for the visibility loss expected from the vFactor (only for SC, that is  $v_{fbl} = 1.0$  for FT).

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	45 of 51

### Visibility amplitude

$$\widetilde{\text{visAmp}}_{bl} = \frac{\sqrt{(\sum_f R'_{fbl})^2 + (\sum_f I'_{fbl})^2}}{\sum_f \sqrt{FF_{fbl} v_{fbl}}} \quad (23)$$

### Visibility phase

$$\widetilde{\text{visPhi}}_{bl} = \arctan\left(\sum_f I'_{fbl}, \sum_f R'_{fbl}\right) \quad (24)$$

Then this phase is cleanup from its mean spectral slope (stored in the GDELAY\_Astrometry quantity) and mean spectral value (stored in the PHASE\_Astrometry quantity).

## D.12 Averaged square visibility estimator

The square visibility of each frames are averaged together accounting for the visibility loss expected from the vFactor (only for SC, that is  $v_{fbl} = 1.0$  for FT).

For a given baseline, the averaging is performed only with accepted frames of this baseline.

$$\widetilde{\text{vis2}}_{bl} = \frac{\sum_f R_{fbl}^2 + \sum_f I_{fbl}^2 - \sum_f \text{var}(R)_{fbl} - \sum_f \text{var}(I)_{fbl}}{\sum_f (FF_{fbl} v_{fbl})} \quad (25)$$

## D.13 Averaged closure-phase estimator

The averaged bispectrum of triplet  $b_{ijk}$  is computed by:

$$\widetilde{B}_{b_{ijk}l} = \sum_f (R_{fb_{ij}l} + i I_{fb_{ij}l}) \cdot (R_{fb_{jk}l} + i I_{fb_{jk}l}) \cdot (R_{fb_{ik}l} - i I_{fb_{ik}l}) \quad (26)$$

For a given triplet, the averaging is performed only with frames for which all three baselines forming the triangles are accepted. Then the closure phase is computed:

$$\widetilde{\text{t3Phi}}_{b_{ijk}l} = \arctan(\widetilde{B}_{b_{ijk}l}) \quad (27)$$

TBD: normalized amplitude of bispectra.

## D.14 Uncertainty on averaged quantities

The uncertainty on averaged quantities is computed by bootstrapping over the accepted frames. Basic fundamentals about this technic can be in the numerical recipes book, section 15.6. See also: <https://en.wikipedia.org/wiki/Bootstrapping>

The implementation of the bootstrapping method makes use of *segmentation* when the number of frame is larger than 100, or of *Monte-Carlo* when the number of frame is smaller than 5.

ESO	GRAVITY Pipeline User Manual	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	46 of 51

### Segmentation (mostly for FT)

When the number of frame is larger than 100, the dataset is first split into a smaller number of *segments* (typically 20 to 100). First integration is done inside the segments. The final variance is estimated by bootstrapping over the segments. The number of segment does change the temporal sampling of the bootstrap, and thus the estimation of the uncertainties. We choose the number of segment so that each is about 1s.

Note that this is mostly relevant for the FT, since the SC has often less than 100 frames (so each frame is a segment).

### Monte-Carlo (mostly for very long DIT on SC)

When the number of frame is lower than 5, the bootstrap method provides unrealistically small uncertainties. Hence the dataset is complemented with few fake frames (up to 5), on which a random realisation of the theoretical noise is added (using the theoretical variance).

This noise is added on the correlated flux quantity only, not on the photometric fluxes, vFactor, reference phase... Moreover it is clear that Monte-Carlo propagates the *fundamental* uncertainties only (photon and detector noise), but not the *atmospheric* noises (effect of injection fluctuation, tracking quality fluctuation).

As a matter of fact, the final uncertainty when the number of frame is very low can still be under-estimated.

## D.15 Processing of MET and FDDL

### From Volts to real time SC-FT phases

The signal of all diode is analysed independently, without prior correction of the voltage, and converted into a real-time SC-FT phase for each diode. This real-time SC-FT phase is unwrapped and forced to match the OCS.MET reference. Both the unwrapping and the absolute reference are done with fringe-integer corrections only. These quantities are labeled OPD\_TEL and OPD\_FC in the OI\_VIS\_MET table.

### Averaging SC-FT phase inside SC DIT

The phase of each diode is averaged inside each SC DIT, as a scalar quantity (not phasor). This averaged signal is stored for each beam in OI\_FLUX table and each base in OI\_VIS table. These quantities are labeled OPD\_MET\_TEL and OPD\_MET\_FC.

### Average the FDDL inside SC DIT

The mean SC and FT FDDLs strain gauge voltage during each frame of the SC, and per beam, is stored in the OI\_FLUX table. Columns are labelled SC\_POS and FT\_POS.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	47 of 51

## D.16 Applying dispersion correction to MET

The DISP\_MODEL provides, for each beam  $t$ :

- $n_{\text{mean}_{tl}}$  : the *mean* refractive index of SC and FT FDDL at wavelength  $l$ , normalize to the one at the metrology wavelength.
- $n_{\text{diff}_{tl}}$  : the *differential* refractive index between SC and FT FDDL wavelength  $l$ , normalize to the one at the metrology wavelength.
- $\alpha_{tm}$  : the non-linearity coefficients of order  $m$  of the SC FDDL.
- $\beta_{tm}$  : the non-linearity coefficients of order  $m$  of the FT FDDL.

### Correction of FDDL non linearity

SC\_POS and FT\_POS are first corrected from non-linearity of the strain gauge, and then co-added:

$$\text{FDDL}_{tf} = \sum_m ( \alpha_{tm} \text{SC\_POS}_{tf}^m + \beta_{tm} \text{FT\_POS}_{tf}^m ) \quad (28)$$

### Dispersive metrology signal

Then the corrected metrology signal is computed with:

$$\text{OPD\_DISP}_{bfl} = n_{\text{mean}_{t_1 l}} \text{OPD\_MET\_FC}_{t_1 f} - n_{\text{mean}_{t_2 l}} \text{OPD\_MET\_FC}_{t_2 f} + \quad (29)$$

$$n_{\text{diff}_{t_1 l}} \text{FDDL}_{t_1 f} / 2 - n_{\text{diff}_{t_2 l}} \text{FDDL}_{t_2 f} / 2 \quad (30)$$

**WARNING:** Note that this dispersive metrology signal contains a fraction of group-delay coded as a spectral slope... hence you should be aware of what are doing with !!

### Dispersive group-delay and remaining phase

This is the group-delay introduced by FDDL in the middle of the band. It is obviously to a constant (the so called zero). This group delay can be used to compute a group-delay astrometry with the SC and FT group-delay.

$$\text{GDELAY\_DISP}_{bf} = \frac{\lambda_{l_1}^{-1} \text{OPD\_DISP}_{l_1 bf} - \lambda_{l_2}^{-1} \text{OPD\_DISP}_{l_2 bf}}{\lambda_{l_1}^{-1} - \lambda_{l_2}^{-1}} \quad (31)$$

where  $l_1$  and  $l_2$  are two consecutive wavelength channel in the middle of the band.

The remaining phase is the following:

$$\text{PHASE\_DISP}_{lbf} = \arctan( \exp( \frac{2i\pi}{\lambda_f} (\text{OPD\_DISP}_{lbf} - \text{GDELAY\_DISP}_{bf}) ) ) \quad (32)$$

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	48 of 51

## E Installation

One is advised to read the installation instructions delivered with the GRAVITY pipeline distribution kit. These release-specific instructions can be found in the file `README` located in the top-level directory of the unpacked GRAVITY pipeline source tree. The supported platforms are listed in Section E.1. It is recommended reading through Section E.2.2 before starting the installation.

A bundled version of the GRAVITY pipeline with all the required tools and an installer script is available from <http://www.eso.org/pipelines/>, for users who are not familiar with the installation of software packages.

### E.1 Supported platforms

The utilisation of the GNU build tools should allow to build and install the GRAVITY pipeline on a variety of UNIX platforms, but it has only been verified on the VLT target platforms:

- Linux (glibc 2.1 or later),
- Sun Solaris 2.8 or later,

using the GNU C compiler (version 3.2 or newer).

### E.2 Building the GRAVITY pipeline

This section shows how to obtain, build and install the GRAVITY pipeline from the official source distribution.

#### E.2.1 Requirements

To compile and install the GRAVITY pipeline one needs:

- the GNU C compiler (version 3.2 or later),
- the GNU `gzip` data compression program,
- a version of the `tar` file-archiving program, and,
- the GNU `make` utility.

An installation of the Common Pipeline library (CPL) must also be available on the system. Currently the CPL version 2.1.1 or newer is required. The CPL distribution can be obtained from [3].

Please note that CPL itself depends on an existing `qfits` installation. The `qfits` sources are available from the CPL download page or directly from the `qfits` homepage at <http://www.eso.org/projects/aot/qfits>. In conjunction with CPL 2.1.1 `qfits` 5.3.1 must be used.



<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	49 of 51

In order to run the GRAVITY pipeline recipes a front-end application is also required. Currently there are two such applications available, a command-line tool called *EsoRex* and the Java based data file organizer, *Gasgano*, which provides an intuitive graphical user interface (see Section 5.2, page 17). At least one of them must be installed. The *EsoRex* and *Gasgano* packages are available at <http://www.eso.org/cpl/esorex.html> and <http://www.eso.org/gasgano> respectively.

For installation instructions of any of the additional packages mentioned before please refer to the documentation of these packages.

## E.2.2 Compiling and installing the GRAVITY pipeline

The GRAVITY pipeline distribution kit 1.0 contains:

gravity-manual-1.0.pdf	The GRAVITY pipeline manual
install_pipeline	Install script
cpl-7.0.tar.gz	CPL 7.0
esorex-3.12.3.tar.gz	esorex 3.12.3
gasgano-2.4.8.tar.gz	GASGANO 2.4.8for Linux
gravity-0.9.5.tar.gz	GRAVITY 0.9.5
gravity-calib-0.9.5.tar.gz	GRAVITY calibration files 0.9.5

Here is a description of the installation procedure:

1. Change directory to where you want to retrieve the GRAVITY pipeline recipes 0.9.5package. It can be any directory of your choice but not:

```
$HOME/gasgano
$HOME/.esorex
```

2. Download from the ESO ftp server, <http://www.eso.org/pipelines/>, the latest release of the GRAVITY pipeline distribution.
3. Verify the checksum value of the tar file with the cksum command.
4. Unpack using the following command:

```
tar -xvf \pipename-kit-\pipelinevers.tar
```

Note that the size of the installed software (including *Gasgano*) together with the static calibration data is about 1.3 Gb, mainly du to the test data.

5. Install: after moving to the top installation directory,

```
cd \pipename-kit-\pipelinevers
```

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	50 of 51

it is possible to perform a simple installation using the available installer script (*recommended*):

```
./install_pipeline
```

(beware: the execution may take a few minutes on Linux and several minutes on SunOS).

Note that this release still needs to link to the eclipse library. At the end of the installation the user in addition to follow what reported by the installation script, needs to source an file (\$HOME/..eclipse\_bash.rc or \$HOME/..eclipse\_bash.rc, depending from the user shell) to set a few environment variables used by a few low level eclipse library based modules.

By default the script will install the GRAVITY recipes, *Gasgano*, *EsoRex*, all the necessary libraries, and the static calibration tables, into a directory tree rooted at \$HOME. A different path may be specified as soon as the script is run.

The only exception to all this is the *Gasgano* tool, that will always be installed under the directory \$HOME/gasgano. Note that the installer will move an existing \$HOME/gasgano directory to \$HOME/gasgano.old before the new *Gasgano* version is installed.

Important: the installation script would ensure that any existing *Gasgano* and *EsoRex* setup would be inherited into the newly installed configuration files (avoiding in this way any conflict with other installed instrument pipelines).

Alternatively, it is possible to perform a manual installation (*experienced users only*): the README file located in the top installation directory contains more detailed information about a step-by-step installation.

<b>ESO</b>	<b>GRAVITY Pipeline User Manual</b>	Doc:	VLT-MAN-ESO-19500-XXXX
		Issue:	Issue 0.2
		Date:	Date October 2016
		Page:	51 of 51

## F Abbreviations and acronyms

ANSI	American National Standards Institute
ASCII	American Standard Code for Information Interchange
CalibDB	Calibration Database
CPL	Common Pipeline Library
DFO	Data Flow Operations department
DFS	Data Flow System department
DHS	Data Handling Server
DO	Data Organizer
DMD	Data Management and Operations Division
DRS	Data Reduction System
ESO	European Southern Observatory
ESOREX	ESO-Recipe Execution tool
FDDL	Fibered Differential Delay Lines
FITS	Flexible Image Transport System
FOV	Field Of View
FT	Fringe-Tracker
SC	Science-Combiner
GUI	Graphical User Interface
OB	Observation Block
OIFITS	OIFITS format, see [4]
PSO	Paranal Science Operations
QC	Quality Control
RON	Read Out Noise
SOF	Set Of Frames
UT	Unit Telescope
AT	Auxiliary Telescope
VLTI	Very Large Telescope Interferometer